

analyzer, and oxides of nitrogen analyzer (certain analyzers may require more frequent calibration depending on particular equipment and uses).

* * * * *

38. Section 86.519-90 of Subpart F is amended by revising paragraphs (d)(1), (d)(4), and (d)(7) to read as follows:

§ 86.519-90 Constant volume sampler calibration.

* * * * *

(d) * * *

(1) Obtain a small cylinder that has been charged with pure propane or carbon monoxide gas (CAUTION—carbon monoxide is poisonous).

* * * * *

(4) Following completion of step (3) above (if methanol injection is required), continue to operate the CVS in the normal manner and release a known quantity of pure methanol (in gaseous form) into the system during the sampling period (approximately 5 minutes). This step does not need to be performed with each verification,

provided that it is performed at least twice annually.

* * * * *

(7) The cause for any discrepancy greater than ± 2 percent must be found and corrected. The Administrator, upon request, may waive the requirement to comply with ± 2 percent methanol recovery tolerance, and instead require compliance with a higher tolerance (not to exceed ± 6 percent), provided that:

(i) The Administrator determines that compliance with the specified tolerance is not practically feasible; and

(ii) The manufacturer makes information available to the Administrator which indicates that the calibration tests and their results are consistent with good laboratory practice, and that the results are consistent with the results of calibration testing conducted by the Administrator.

39. Section 86.521-90 of Subpart F is amended by revising paragraphs (d) introductory text, (d)(1), and (d)(3)(iii) to read as follows:

§ 86.521-90 Hydrocarbon analyzer calibration.

* * * * *

(d) *FID response factor to methanol.*
When the FID analyzer is to be used for the analysis of hydrocarbon samples containing methanol, the methanol response factor of the analyzer shall be established. The methanol response factor shall be determined at several concentrations in the range of concentrations in the exhaust sample, using either bag samples or gas bottles meeting the requirements of § 86.514.

(1) The bag sample, if used, of methanol for analysis in the FID shall be prepared using the apparatus shown in Figure F90-4. A known volume of methanol is injected, using a microliter syringe, into the heated mixing zone (250°F (121°C)) of the apparatus. The methanol is vaporized and swept into the sample bag with a known volume of zero grade air measured by a gas flow meter meeting the performance requirements of § 86.120.

BILLING CODE 6560-50-P